



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

ETS-LINDGREN INC.
1301 Arrow Point Drive
Cedar Park, TX 78613
Ron Bethel Phone: 512 531 6400

CALIBRATION

Valid To: March 31, 2019

Certificate Number: 1207.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – RF/Microwave

Parameter/Equipment	Frequency	CMC ² (±)	Comments
Conical Log Spiral Antenna – Antenna Factor	(100 to 300) MHz (300 to 850) MHz (850 to 1000) MHz (1 to 10) GHz	1.4 dB 1.2 dB 0.93 dB 1.2 dB	SAE/ARP 958D
Biconical Antenna – Antenna Factor	(20 to 30) MHz (30 to 2000) MHz (2 to 18) GHz (20 to 300) MHz (300 to 2000) MHz (2 to 8) GHz (8 to 18) GHz	1.1 dB 1.7 dB 2.5 dB 0.9 dB 1.6 dB 2.0 dB 3.6 dB	SAE/ARP 958D ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)

Parameter/Equipment	Frequency	CMC ² (±)	Comments
Ridge Guide Horn Antenna – Antenna Factor	(0.2 to 2) GHz (2 to 8) GHz (8 to 18) GHz (18 to 40) GHz (0.2 to 18) GHz	1.7 dB 1.3 dB 1.7 dB 1.3 dB 1 dB	SAE/ARP 958D ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)
Microwave Horn Antenna – Antenna Factor	(0.7 to 2) GHz (2 to 8) GHz (8 to 18) GHz (18 to 40) GHz (1 to 40) GHz	0.78 dB 0.93 dB 0.97 dB 1.3 dB 0.5 dB	SAE/ARP 958D ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)
Dipole Antennas – Antenna Factor	(30 to 60) MHz (60 to 140) MHz (140 to 400) MHz (400 to 1000) MHz	0.9 dB 0.7 dB 0.8 dB 1 dB	ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)
Fixed Length Dipole – Antenna Factor	(420 to 3010) MHz	1.3 dB	1150-3-3012 (three antenna method in fully anechoic chamber)
Sleeve Dipole and Resonant Loop, Gain Efficiency	400 MHz to 6 GHz	0.19 dB	1650-3-0007 (at resonant frequency only)
Biconilog Antennas – Antenna Factor	(26 to 30) MHz 30 MHz to 1 GHz (1 to 7) GHz (26 to 30) MHz 30 MHz to 1 GHz (1 to 7) GHz	1.0 dB 1.3 dB 2.1 dB 1.5 dB 1 dB 1.4 dB	SAE/ARP 958D ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)

Parameter/Equipment	Frequency	CMC ² (±)	Comments
Log-Periodic Antennas – Antenna Factor	(200 to 2) GHz	1.0 dB	SAE/ARP 958D
	(26 to 30) MHz	1.4 dB	ANSI C63.5:1988, ANSI C63.5:1998, ANSI C63.5:2006 ANSI C63.5:2017 (standard site method)
	30 MHz to 1 GHz	0.8 dB	
	(1 to 5) GHz	1.2 dB	
(5 to 18 GHz)	1.8 dB		
Antenna VSWR	(10 to 100) MHz	1.4 dB	1150-3-3020
	100 MHz to 6 GHz	0.33 dB	
Antenna Symmetry	(30 to 1000) MHz	1.7 dB	CISPR 16-1-4
	(30 to 300) MHz	1.7 dB	ANSI C63.5
	20 MHz to 2 GHz	1.7 dB	1150-3-3020
	(2 to 6) GHz	2 dB	

II. Electrical – RF/EMC Instrument Parameters

Parameter/Equipment	Frequency	CMC ² (±)	Comments
Monopole/Rod Antenna – Antenna Factor	30 Hz to 50 MHz	1.8 dB	IEEE/IEC 291, CISPR 16-1-4 (2010) ANSI C63.5 (ECSM) SAE/ARP 958D
Loop Antenna – Antenna Factor	20 Hz to 30 MHz	0.89 dB	IEEE/IEC 291 ANSI C63.5:2017 (thermocouple method)

Parameter/Equipment	Frequency	CMC ² (±)	Comments
E-Field Probe – Frequency Response, Linearity	10 kHz to 18 GHz (26.5 to 40) GHz	0.64 dB 0.95 dB	IEEE 1309 (TEM cell, GTEM, anechoic test chamber)
Frequency Response, Linearity	80 MHz to 6 GHz	7.7 % of reading	IEC 61000-4-3
Isotropy	80 MHz to 6 GHz	0.86 dB	IEEE 1309
Magnetic Field Meter, Frequency Response, Linearity	50 kHz to 220 MHz	1.2 dB	IEEE 1309 (standard field method)
LISN (Line Impedance Stabilization Network) – Insertion Loss	9 kHz to 100 MHz	0.44 dB	ANSI C63.4, CISPR 25, CISPR 16-1-2
Impedance – Magnitude	(9 to 25) kHz (25 to 100) kHz 100 kHz to 1.5 MHz (1.5 to 100) MHz	2.0 % of reading 2.0 % of reading 2.0 % of reading 2.0 % of reading	
Impedance – Phase	9 kHz to 100 MHz	3.8°	CISPR 16-1-2
Isolation	9 kHz to 100 MHz	0.47 dB	CISPR 16-1-2
PLISN (Power Line Impedance Stabilization Network) – Insertion Loss	9 kHz to 1 GHz	0.84 dB	1150-3-3054
Impedance – Magnitude	(9 to 25) kHz (25 to 100) kHz 100 kHz to 1.5 MHz (1.5 to 100) MHz 100 MHz to 1 GHz	1.9 dB 1.4 dB 1.3 dB 1.4 dB 1.4 dB	

Parameter/Equipment	Frequency	CMC ² (±)	Comments
Line Impedance Probe – Insertion Loss Impedance – Magnitude	9 kHz to 100 MHz (9 to 100) kHz 100 kHz to 10 MHz (10 to 100) MHz	0.99 dB 2.8 % of rdg 2.3 % of rdg 5.5 % of rdg	1150-3-3021
RF Cable – Insertion Loss	10 Hz to 3 GHz (3 to 40) GHz	0.61 dB 1.2 dB	1150-3-3040
RF Directional Couplers – Insertion Loss	9 kHz to 3 GHz <40 dB <55 dB <70 dB (3 to 40) GHz <40 dB <55 dB <70 dB	1.2 dB 1.3 dB 1.7 dB 2.1 dB 2.4 dB 5.5 dB	1150-3-3025
RF Attenuators – Attenuation	10 Hz to 3 GHz <40 dB <55 dB <70 dB (3 to 40) GHz <40 dB <55 dB <70 dB	0.62 dB 0.66 dB 1.4 dB 1.7 dB 2.1 dB 5.3 dB	1150-3-3041
Current Probe – Transfer Impedance Insertion Loss	10 Hz to 3 GHz 10 Hz to 3 GHz	2.3 dB 2.0 dB	CISPR 16-1-2

Parameter/Equipment	Frequency	CMC ² (±)	Comments
RF Power Sensor –			
Frequency Response	9 kHz to 6 GHz (6 to 18) GHz	0.60 dB 0.61 dB	1150-3-3067
VSWR	(10 to 100) MHz 100 MHz to 2 GHz (2 to 6) GHz	0.28 dB 0.28 dB 0.28 dB	
Linearity	(50 and 500) MHz	0.86 dB	
Linearity (-45 dBm)	(50 and 500) MHz	0.93 dB	
RF Signal Generator –			
Frequency Response	9 kHz to 6 GHz	0.63 dB	1150-3-3064

III. Electrical – Field Calibrations

Equipment Type	Frequency	CMC ² (±)	Comments
Wireless Test Chamber Calibrations ³ –			
Range Calibration	<6 GHz	1.3 dB	CTIA test plan for wireless device over-the-air performance
Ripple Test	<6 GHz	1.3 dB	
Normalized Site Attenuation ³ (OATS)	30 MHz to 1 GHz	1.7 dB	ANSI C63.4
Normalized Alternative Site Attenuation ³	30 MHz to 1 GHz	1.7 dB	ANSI C63.4
Field Uniformity ³	26 MHz to 6 GHz	1.5 dB 1.4 dB	IEC 61000-4-3 IEC 61000-4-21
SVSWR ³	(1 to 18) GHz	1.3 dB	CISPR 16-1-4

¹ This laboratory offers commercial calibration service and field calibration service, where noted.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



Accredited Laboratory

A2LA has accredited

ETS-LINDGREN INC.

Cedar Park, TX

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 7th day of August 2017.

A handwritten signature in blue ink, appearing to read "L. Anderson", written over a horizontal line.

President & CEO
For the Accreditation Council
Certificate Number 1207.01
Valid to March 31, 2019
Revised on December 28, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.