



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540.3

GENERAL CALIBRATION, A TRESICAL COMPANY
 2 Mars Court
 Boonton, NJ 07005
 Deborah McLaren Phone: 973 299 2950

CALIBRATION

Valid To: January 31, 2018

Certificate Number: 2353.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Chemical

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
pH – Measuring Equipment ³	(4, 7, 10) pH	0.016 pH	Buffer solutions
Electrolytic Conductivity – Measuring Equipment ³	≈10 μS/cm ≈100 μS/cm ≈1000 μS/cm ≈10 000 μS/cm	0.53 μS/cm 2.4 μS/cm 23 μS/cm 0.23 mS/cm	Conductivity solutions

II. Dimensional

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Flatness ³	Up to 1 in	4.6 μin	Monochromatic light and optical flats
Hand Tools ³ Indicators Micrometers Depth Gages Height Gages Calipers	Up to 1 in	25 μin /in + 0.8 μin	Gage blocks

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Pin and Plug Gages	(0.01 to 1) in	120 µin/in	Gage blocks and micrometer
Protractors and Tilt Sensors	Up to 60°	5.8"	Gage blocks & sine bar

III. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4,5,6} (±)	Comments
DC Voltage – Generate	Up to 220 mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	7.1 µV/V + 0.40 µV 3.7 µV/V + 0.80 µV 4.9 µV/V + 3.0 µV 4.9 µV/V + 4.3 µV 3.7 µV/V + 48 µV 4.7 µV/V + 0.48 mV	Fluke 5720A
Fixed Value	10.0 V	3.4 µV/V	Fluke 732A
DC Voltage ³ – Measure	(0 to 0.1) V (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1100) V	9.5 µV/V + 0.37 µV 6.1 µV/V + 0.37 µV 5.8 µV/V + 0.60 µV 8.7 µV/V + 37 µV 8.7 µV/V + 0.12 mV	Agilent 3458A
DC High Voltage ³ – Generate & Measure	(1 to 45) kV (45 to 200) kV	0.11 % 2.4 %	VD45 Ross divider VMP200 Ross divider
DC Current ³ – Generate	Up to 220 µA (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A (2.2 to 11) A (0 to 10.9999) A (11 to 20.5) A (20 to 150) A (150 to 1000) A	35 µA/A + 6.0 nA 31 µA/A + 7.0 nA 30 µA/A + 41 nA 41 µA/A + 0.71 µA 77 µA/A + 12 µA 0.034 % + 0.48 mA 0.057 % + 500 µA 0.11 % + 750 µA 0.28 % + 500 µA 0.57 % + 500 µA	Fluke 5720A Fluke 5520A Fluke 5500A Coil/5520A

Parameter/Equipment	Range	CMC ^{2, 4, 5, 6} (±)	Comments
DC Current ³ – Measure	(0 to 100) μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	17 μ A/A + 0.80 nA 17 μ A/A + 5.0 nA 17 μ A/A + 50 nA 31 μ A/A + 0.50 μ A 0.010 % + 10 μ A	Agilent 3458A
Resistance ³ – Generate	Up to 10.9999 Ω (11 to 32.9999) Ω (33 to 109.9999) Ω (110 to 329.9999) Ω (0.33 to 1.099999) k Ω (1.1 to 3.299999) k Ω (3.3 to 10.99999) k Ω (11 to 32.99999) k Ω (33 to 109.999) k Ω (110 to 329.999) k Ω	40 $\mu\Omega/\Omega$ + 1.0 m Ω 30 $\mu\Omega/\Omega$ + 2.0 m Ω 28 $\mu\Omega/\Omega$ + 2.0 m Ω 28 $\mu\Omega/\Omega$ + 4.0 m Ω 28 $\mu\Omega/\Omega$ + 13 m Ω 28 $\mu\Omega/\Omega$ + 13 m Ω 28 $\mu\Omega/\Omega$ + 30 m Ω 28 $\mu\Omega/\Omega$ + 0.30 Ω 28 $\mu\Omega/\Omega$ + 0.30 Ω 32 $\mu\Omega/\Omega$ + 2.0 Ω	Fluke 5520A, 4-wire
	(0.33 to 1.09999) M Ω (1.1 to 3.29900) M Ω (3.3 to 10.9999) M Ω (11 to 32.9999) M Ω (33 to 109.9999) M Ω (110 to 329.9999) M Ω (330 to 1100) M Ω	32 $\mu\Omega/\Omega$ + 2.2 Ω 60 $\mu\Omega/\Omega$ + 39 Ω 0.013 % + 63 Ω 0.025 % + 2.5 k Ω 0.050 % + 3.0 k Ω 0.30 % + 0.10 M Ω 1.5 % + 0.50 M Ω	Fluke 5520A, 2-wire
Fixed Points	0 Ω 1 Ω 1.9 Ω 10, 19 Ω 100, 190 Ω 1, 1.9 k Ω 10, 19 k Ω 100, 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	41 $\mu\Omega$ 81 $\mu\Omega/\Omega$ 82 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 9.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 30 $\mu\Omega/\Omega$ 35 $\mu\Omega/\Omega$ 45 $\mu\Omega/\Omega$ 0.010 %	Fluke 5720A
	1 M Ω 10 M Ω 100 M Ω	0.11 % per decade 0.11 % per decade 0.23 % per decade	IET HRRS-B-3-1M

Parameter/Equipment	Range	CMC ^{2, 4, 5, 6} (±)	Comments
Resistance ³ – Measure	100 mΩ to 10 Ω (10 to 100) Ω (0.1 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ (0.1 to 1) GΩ	18 μΩ/Ω + 58 μΩ 12 μΩ/Ω + 0.58 mΩ 9.5 μΩ/Ω + 5.8 mΩ 9.5 μΩ/Ω + 58 mΩ 9.7 μΩ/Ω + 0.58 Ω 15 μΩ/Ω + 2.3 Ω 58 μΩ/Ω + 0.12 kΩ 0.058 % + 1.2 kΩ 0.59 % + 12 kΩ	Agilent 3458A
Shunt ³ – AC/DC, Fixed Point	500.78 μΩ	0.028 %	ITT HA0511 100A, 50 mV
Ground Bond Test ³ – AC/DC, Fixed Point	0.1 Ω	33 μΩ/Ω	Ground bond test resistor
Corona Simulation ³	(1 to 1000) pC	4 %	Agilent 34401A w/ Andeen-Hagerling AH2550

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Voltage ³ – Generate Up to 2.2 mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 4.0 μV 85 μV/V + 4.0 μV 75 μV/V + 4.0 μV 0.018 % + 4.0 μV 0.046 % + 5.0 μV 0.090 % + 10 μV 0.12 % + 20 μV 0.25 % + 20 μV	Fluke 5720A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 4.0 μV 85 μV/V + 4.0 μV 75 μV/V + 4.0 μV 0.018 % + 4.0 μV 0.046 % + 5.0 μV 0.090 % + 10 μV 0.12 % + 20 μV 0.25 % + 20 μV	

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Voltage ³ – Generate (cont)			
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 12 µV 85 µV/V + 7.0 µV 75 µV/V + 7.0 µV 0.018 % + 7.0 µV 0.042 % + 17 µV 0.075 % + 20 µV 0.12 % + 25 µV 0.25 % + 45 µV	Fluke 5720A
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 82 µV 85 µV/V + 82 µV 40 µV/V + 82 µV 70 µV/V + 82 µV 0.011 % + 82 µV 0.034 % + 82 µV 0.090 % + 0.20 mV 0.15 % + 0.32 mV	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 0.40 mV 80 µV/V + 0.15 mV 40 µV/V + 50 µV 70 µV/V + 0.10 mV 95 µV/V + 0.20 mV 0.026 % + 0.60 mV 0.090 % + 2.0 mV 0.13 % + 3.2 mV	
(22 to 220) V*	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.022 % + 4.0 mV 80 µV/V + 1.5 mV 47 µV/V + 0.61 mV 75 µV/V + 1.0 mV 0.013 % + 2.5 mV 0.080 % + 16 mV 0.42 % + 40 mV 0.70 % + 80 mV	* 220 V range subject to 2.2E7 V-Hz limitation
(220 to 1100) V	40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	80 µV/V + 4.1 mV 0.013 % + 6.1 mV 0.036 % + 11 mV	
(220 to 750) V	(30 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (500 to 1000) KHz	0.036 % + 11 mV 0.080 % + 45 mV 0.13 % + 83 mV 0.42 % + 91 mV 0.70 % + 1.1 V	

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Voltage ³ – Measure			
(0.1 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.030 % + 3.3 μV 0.020 % + 1.8 μV 0.030 % + 1.8 μV 0.10 % + 1.8 μV 0.50 % + 1.8 μV 4.0 % + 1.8 μV	Agilent 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz (1 to 2) MHz	0.007 % + 10 μV 0.007 % + 2 μV 0.014 % + 2 μV 0.030 % + 2 μV 0.080 % + 2 μV 0.30 % + 10 μV 1.0 % + 10 μV 1.5 % + 10 μV	
100 mV to 1V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz (1 to 2) MHz	0.007 % + 40 μV 0.007 % + 20 μV 0.014 % + 20 μV 0.030 % + 20 μV 0.080 % + 20 μV 0.30 % + 100 μV 1.0 % + 100 μV 1.5 % + 100 μV	
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz (1 to 2) MHz	0.007 % + 0.40 mV 0.007 % + 0.20 mV 0.014 % + 0.20 mV 0.030 % + 0.20 mV 0.080 % + 0.20 mV 0.30 % + 1.0 mV 1.0 % + 1.0 mV 1.5 % + 1.0 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.02 % + 4.0 mV 0.02 % + 2.0 mV 0.02 % + 2.0 mV 0.04 % + 2.0 mV 0.12 % + 2.0 mV 0.40 % + 10 mV 1.5 % + 10 mV	

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Voltage ³ – Measure (cont.) (100 to 707) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.04 % + 40 mV 0.04 % + 20 mV 0.06 % + 20 mV 0.12 % + 20 mV 0.30 % + 20 mV	Agilent 3458A
AC High Voltage ³ – Measure (1 to 30) kVrms (30 to 142) kVrms	60 Hz 60 Hz	0.26 % 4.4 %	Ross VD45 Ross VMP200
AC Current ³ – Generate Up to 220 µA (0.22 to 2.2) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 16 nA 0.014 % + 10 nA 0.011 % + 8.0 nA 0.025 % + 12 nA 0.090 % + 65 nA 1.6 % + 0.40 µA 0.023 % + 41 nA 0.014 % + 36 nA 0.011 % + 36 nA 0.025 % + 0.11 µA 0.090 % + 0.65 µA 1.0 % + 0.60 µA	Fluke 5720A w/ 5725A Fluke 5520A Fluke 5720A w/ 5725A Fluke 5520A

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Current ³ – Generate (cont)			
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 0.40 µA 0.014 % + 0.36 µA 0.011 % + 0.36 µA 0.025 % + 0.56 µA 0.090 % + 5.0 µA 0.40 % + 4.0 µA	Fluke 5720A w/ 5725A Fluke 5520A
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 4.1 µA 0.014 % + 3.6 µA 0.011 % + 2.6 µA 0.018 % + 3.6 µA 0.090 % + 10 µA 0.40 % + 0.20 mA	Fluke 5720A w/ 5725A Fluke 5520A
Up to 220 µA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 16 nA 0.014 % + 10 nA 0.011 % + 8.0 nA 0.025 % + 12 nA 0.090 % + 65 nA 1.6 % + 0.40 µA	Fluke 5720A w/ 5725A Fluke 5520A
(0.22 to 2.2) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 41 nA 0.014 % + 36 nA 0.011 % + 36 nA 0.025 % + 0.11 µA 0.090 % + 0.65 µA 1.0 % + 0.60 µA	Fluke 5720A w/ 5725A Fluke 5520A
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 0.40 µA 0.014 % + 0.36 µA 0.011 % + 0.36 µA 0.025 % + 0.56 µA 0.090 % + 5.0 µA 0.40 % + 4.0 µA	Fluke 5720A w/ 5725A Fluke 5520A
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.023 % + 4.1 µA 0.014 % + 3.6 µA 0.011 % + 2.6 µA 0.018 % + 3.6 µA 0.090 % + 10 µA 0.40 % + 0.20 mA	Fluke 5720A w/ 5725A Fluke 5520A

Parameter/Range	Frequency	CMC ^{2, 4, 5, 6} (±)	Comments
AC Current ³ – Generate (cont.)			
(0.22 to 2.2) A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.024 % + 36 µA 0.039 % + 80 µA 0.60 % + 0.16 mA	Fluke 5720A w/ 5725A
(2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.040 % + 0.18 mA 0.085 % + 0.39 mA 0.33 % + 0.75 mA	Fluke 5720A w/ 5725A
(11 to 20.5) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.12 % + 5.1 mA 0.15 % + 5.1 mA 3.0 % + 5.1 mA	Fluke 5520A
(16.5 to 150) A	(45 to 65) Hz (65 to 440) Hz	0.38 % + 0.029 A 1.0 % + 0.031 A	Fluke 5520A w/ coil
(150 to 1025) A	(45 to 65) Hz (65 to 440) Hz	1.0 % + 0.031 A 1.0 % + 0.12 A	
AC Current ³ – Measure			
(5 to 100) µA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.46 % + 0.035 µA 0.18 % + 0.035 µA 0.07 % + 0.035 µA	Agilent 3458A
(0.1 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.46 % + 0.024 % 0.18 % + 0.024 % 0.07 % + 0.024 % 0.036 % + 0.024 %	
(0.1 to 1) A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.46 % + 0.23 mA 0.19 % + 0.23 mA 0.093 % + 0.23 mA 0.12 % + 0.23 mA	
100 mA to 20 A	Up to 1 kHz (1 to 5) kHz	0.039 % + 0.032 %* <i>F</i> 0.041 % + 0.032 %* <i>F</i>	w/ Fluke Y5020 shunt * <i>F</i> is the applied frequency

Parameter/Range	Frequency	CMC ^{2,4,5,6} (±)	Comments	
Capacitance ³ – Generate				
(0.10 to 3.299) nF	10 Hz to 10 kHz	0.51 % + 12 pF	Fluke 5520A	
(0.33 to 10.999) nF	(10 to 1000) Hz	0.26 % + 12 pF		
(11 to 109.999) nF	(10 to 1000) Hz	0.26 % + 0.12 nF		
(110 to 329.99) nF	(10 to 1000) Hz	0.26 % + 0.31 nF		
(0.33 to 1.0999) μF	(10 to 600) Hz	0.26 % + 1.2 nF		
(1.1 to 3.2999) μF	(10 to 300) Hz	0.26 % + 3.1 nF		
(3.3 to 10.999) μF	(10 to 150) Hz	0.26 % + 12 nF		
(11 to 32.999) μF	(10 to 120) Hz	0.42 % + 31 nF		
(33 to 109.99) μF	(10 to 80) Hz	0.46 % + 0.12 μF		
(110 to 329.99) μF	Up to 50 Hz	0.46 % + 0.31 μF		
(0.33 to 1.0999) mF	Up to 20 Hz	0.46 % + 1.2 μF		
(1.1 to 3.2999) mF	Up to 6 Hz	0.46 % + 3.1 μF		
(3.3 to 10.999) mF	Up to 2 Hz	0.46 % + 12 μF		
(11 to 32.999) mF	Up to 0.6 Hz	0.78 % + 31 μF		
(33 to 110) mF	Up to 0.2 Hz	1.2 % + 0.12 mF		
Fixed Points				
1 pF	1 kHz to 13 MHz	0.37 %		Agilent 16381A
10 pF	1 kHz to 13 MHz	0.040 %	Agilent 16382A	
100 pF	1 kHz to 13 MHz	0.067 %	Agilent 16383A	
1000 pF	1 kHz to 13 MHz	0.33 %	Agilent 16384A	
(10, 100, 1000) nF	120 Hz to 100 kHz	0.010 %	Agilent 16380C	

Parameter/Range	Frequency	CMC ^{2,4,5,6} (±)	Comments
Capacitance ³ – Measure			
1 pF	500 Hz to 5 kHz (5 to 100) kHz	12 % 1.2 %	Fluke PM6304C
10 pF	(150 to 500) Hz 500 Hz to 5 kHz (5 to 20) kHz (20 to 100) kHz	12 % 1.2 % 0.12 % 0.46 %	
100 pF	(50 to 250) Hz 250 Hz to 1 kHz (1 to 20) kHz (20 to 100) kHz	12 % 1.2 % 0.12 % 0.46 %	
1 nF	(50 to 250) Hz 250 Hz to 20 kHz (20 to 100) kHz	1.2 % 0.12 % 0.46 %	
10 nF	(50 to 500) Hz 500 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	0.12 % 0.06 % 0.12 % 0.46 %	
100 nF	(50 to 150) Hz 150 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	0.12 % 0.06 % 0.12 % 0.46 %	
1 μF	50 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	0.06 % 0.12 % 0.46 %	
10 μF	(50 to 1500) Hz (1.5 to 15) kHz (15 to 50) kHz (50 to 100) kHz	0.06 % 0.12 % 1.2 % 12 %	
100 μF	(50 to 1500) Hz (1.5 to 15) kHz (15 to 50) kHz	0.12 % 1.2 % 12 %	

Parameter/Range	Frequency	CMC ^{2,4,5,6} (±)	Comments
Inductance ³ – Generate			
Fixed Points			
100 µH	400 Hz & 1 kHz	1.2 %	GenRad 1482 series
1 mH		0.13 %	
10 mH		0.083 %	
100 mH		0.083 %	
1 H		0.083 %	
Inductance ³ – Measure			
1 µH	500 Hz to 1 kHz (50 to 100) kHz	12 % 1.2 %	Fluke PM6304C
10 µH	(250 to 500) Hz 500 Hz to 20 kHz (20 to 100) kHz	12 % 1.2 % 0.46 %	
100 µH	(75 to 250) Hz (250 to 1500) Hz (1.5 to 20) kHz (20 to 100) kHz	12 % 1.2 % 0.12 % 0.46 %	
1 mH	(50 to 75) Hz (75 to 250) Hz 250 Hz to 20 kHz (20 to 100) kHz	12 % 1.2 % 0.12 % 0.46 %	
10 mH	(50 to 75) Hz (75 to 250) Hz 250 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	1.2 % 0.12 % 0.06 % 0.12 % 0.46 %	
100 mH	(50 to 75) Hz 75 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	0.12 % 0.06 % 0.12 % 0.46 %	
1 H	50 Hz to 2 kHz (2 to 20) kHz (20 to 100) kHz	0.06 % 0.12 % 0.46 %	
Capacitance Dissipation Factor ³ –	(0.2 to 5) %	0.25 %	

Parameter/Range	Frequency	CMC ^{2,4,5,6} (±)	Comments
Inductance ³ – Measure (cont)			
10 H	(50 to 250) Hz 250 Hz to 20 kHz (20 to 100) kHz	0.06 % 0.12 % 1.2 %	Fluke PM6304C
100 H	(50 to 2500) Hz (2.5 to 15) kHz (15 to 75) kHz	0.12 % 1.2 % 12 %	
1000 H	(50 to 250) Hz 250 Hz to 2.5 kHz (2.5 to 15) kHz	0.12 % 1.2 % 13 %	
Thermocouple ³ – Indicating Systems & Measure			
Type B	(600 to 800) °C (800 to 1550) °C (1550 to 1820) °C	0.35 °C 0.28 °C 0.22 °C	Fluke 7526A
Type C	(0 to 1000) °C (1000 to 1800) °C (1800 to 2000) °C (2000 to 2316) °C	0.16 °C 0.23 °C 0.26 °C 0.35 °C	
Type E	(-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 600) °C (600 to 1000) °C	0.25 °C 0.12 °C 0.09 °C 0.08 °C 0.1 °C	
Type J	(-210 to -100) °C (-100 to 800) °C (800 to 1200) °C	0.14 °C 0.09 °C 0.1 °C	
Type K	(-250 to -200) °C (-200 to -100) °C (-100 to 500) °C (500 to 800) °C (800 to 1372) °C	0.46 °C 0.16 °C 0.1 °C 0.1 °C 0.13 °C	

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Thermocouple ³ – Indicating Systems & Measure			
Type L	(-200 to -100) °C (-100 to 900) °C	0.1 °C 0.09 °C	Fluke 7526A
Type N	(-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 100) °C (100 to 800) °C (800 to 1300) °C	0.73 °C 0.23 °C 0.12 °C 0.11 °C 0.1 °C 0.12 °C	
Type R	(-50 to -25) °C (-25 to 0) °C (0 to 100) °C (100 to 400) °C (400 to 600) °C (600 to 1000) °C (1000 to 1600) °C (1600 to 1767) °C	0.55 °C 0.45 °C 0.39 °C 0.28 °C 0.22 °C 0.21 °C 0.19 °C 0.23 °C	
Type S	(-50 to -25) °C (-25 to 0) °C (0 to 100) °C (100 to 400) °C (400 to 600) °C (600 to 1000) °C (1000 to 1600) °C (1600 to 1767) °C	0.51 °C 0.43 °C 0.38 °C 0.29 °C 0.23 °C 0.22 °C 0.22 °C 0.26 °C	
Type T	(-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 200) °C (200 to 400) °C	0.35 °C 0.16 °C 0.11 °C 0.09 °C 0.09 °C	
Type U	(-200 to 0) °C (0 to 200) °C (200 to 600) °C	0.16 °C 0.1 °C 0.1 °C	

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Electrical Calibration of RTDs ³ – Generate			
Pt 385, 100 Ω	(-200 to -80) °C (-80 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 800) °C	0.013 °C 0.02 °C 0.024 °C 0.026 °C 0.033 °C 0.038 °C	Fluke 7526A
Pt 3926, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.013 °C 0.015 °C 0.017 °C 0.022 °C 0.026 °C 0.032 °C	
Pt 3916, 100 Ω	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.01 °C 0.013 °C 0.015 °C 0.017 °C 0.022 °C 0.026 °C 0.031 °C 0.033 °C	
Pt 385, 200 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 630) °C	0.053 °C 0.056 °C 0.06 °C 0.06 °C 0.069 °C 0.071 °C 0.088 °C	
Pt 385, 500 Ω	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.026 °C 0.028 °C 0.034 °C 0.038 °C 0.045 °C	
Pt 385, 1000 Ω	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.015 °C 0.018 °C 0.024 °C 0.026 °C 0.033 °C	

Parameter/Equipment	Range	CMC ^{2,4,5,6} (±)	Comments
Electrical Calibration of RTDs ³ – Generate (cont)			
Ni 385, 120 Ω	(-80 to 260) °C	0.009 °C	Fluke 7526A
Cu 427, 10 Ω	(-100 to 260) °C	0.11 °C	
SPRT	(-200 to 660) °C	0.06 °C	
Power Measuring Equipment AC, DC Power ³ – Generate			
PF = 1: (33 to 329.99) mV (45 to 65) Hz	(3.3 to 8.999) mA (9 to 32.999) mA (33 to 89.999) mA (90 to 329.99) mA (0.33 to 0.8999) A (0.9 to 2.1999) A (2.2 to 4.4999) A (4.5 to 20.5) A	0.16 % 0.11 % 0.16 % 0.12 % 0.15 % 0.12 % 0.15 % 0.13 %	Fluke 5520A
330 mV to 1020 V (45 to 65) Hz	(3.3 to 8.999) mA (9 to 32.999) mA (33 to 89.999) mA (90 to 329.99) mA (0.33 to 0.8999) A (0.9 to 2.1999) A (2.2 to 4.4999) A (4.5 to 20.5) A	0.14 % 0.09 % 0.14 % 0.09 % 0.13 % 0.1 % 0.14 % 0.12 %	
Phase/Power Factor ³ – (10 to 65) Hz PF (0 to 1)	0Φ / PF 1 10Φ / PF 0.985 20Φ / PF 0.940 30Φ / PF 0.866 40Φ / PF 0.766 50Φ / PF 0.643 60Φ / PF 0.500 70Φ / PF 0.342 80Φ / PF 0.174	0.58 % 0.29 % 0.22 % 0.22 % 0.26 % 0.35 % 0.55 % 1.1 %	Fluke 5520A

Parameter/Equipment	Range	CMC ^{2, 4, 6} (±)	Comments
Phase Angle ³ – Measure (0 to 359.9)°	20 Hz to 10 kHz (10 to 40) kHz (40 to 100) kHz	0.081 ° 0.29 ° 0.98 °	Krohn-Hite 6500
Oscilloscopes ³ – Rise Time Bandwidth	Single Sided 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (0.6 to 1.1) GHz (1.1 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	< 300 ps ± 120 ps 3.5 % + 300 μV 4 % + 300 μV 6 % + 300 μV 7 % + 300 μV 0.31 dB 0.74 dB 0.84 dB	Fluke 5520A-SC1100 Agilent 8340A w/: 8482A, 11667A 8481A, 11667A 8485A, 11667B
Audio Distortion ³ (THD)	Up to 100 kHz	1.2 dB	Agilent 8903B

IV. Electrical – RF/Microwave

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
Power Meter – Power Reference, @ 1 mW	50 MHz	1.9 %	Agilent 432A w/ 478A
Relative Power (Tuned RF Level) – Measure 100 kHz to 50 GHz	(0 to -10) dB (-10 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (-50 to -60) dB (-60 to -70) dB (-70 to -80) dB (-80 to -90) dB	0.018 dB 0.019 dB 0.019 dB 0.056 dB 0.056 dB 0.057 dB 0.057 dB 0.094 dB 0.094 dB	Agilent E4448A

Parameter/Range	Frequency	CMC ^{2, 5, 6} (\pm)	Comments
Relative Power (Tuned RF Level) – Measure (cont) 100 kHz to 50 GHz	(-90 to -100) dB (-100 to -110) dB (-110 to -120) dB (-120 to -130) dB	0.094 dB 0.095 dB 0.096 dB 0.096 dB 0.097 dB	Agilent E4448A
Absolute Power – Measure (-30 to -50) dBm (-50 to -60) dBm (-60 to 68) dBm (-30 to +10) dBm (+10 to +20) dBm	10 MHz to 18 GHz 100 kHz to 4.2 GHz (4.2 to 18) GHz (18 to 26.5) GHz (26.5 to 50) GHz 100 kHz to 4.2 GHz (4.2 to 18) GHz (18 to 26.5) GHz (26.5 to 50) GHz	1.3 % 2.6 % 15 % 3.3 % 3.5 % 4.2 % 5.0 % 5.9 % 6.0 % 6.5 % 7.0 %	Agilent N5531S w/ N-Type power sensor, w/ power meter 3.5 mm 2.4 mm 3.5 mm 2.4 mm
VSWR	5 MHz to 2 GHz (2 to 12.5) GHz (12.5 to 18) GHz	0.11 dB 0.53 dB 0.85 dB	E4448A opt 233 Measuring receiver w/ SWR bridges
Frequency Modulation – Measure Mod Rate: 20 Hz to 10 kHz Dev.: 200 Hz to 40 kHz ($\beta > 0.2$) Mod Rate: 50 Hz to 200 kHz Dev.: 250 Hz to 400 kHz ($\beta > 0.2$)	250 kHz to 10 MHz 10 MHz to 6.6 GHz (6.6 to 13.2) GHz (13.2 to 31.15) GHz (31.15 to 50) GHz	1.0 % 1.0 % 1.0 % 1.0 % 1.0 %	E4448A opt 233 β is the ratio of the frequency deviation to the modulation rate

Parameter/Range	Frequency	CMC ^{2, 5, 6} (\pm)	Comments
Amplitude Modulation – Measure Depth: (5 to 99) % (5 to 20) % (20 to 99) % (5 to 20) % (20 to 99) % (5 to 20) % (20 to 99) % (5 to 20) % (20 to 99) %	100 kHz to 10 MHz 10 MHz to 3 GHz 10 MHz to 3 GHz (3 to 26.5) GHz (3 to 26.5) GHz (26.5 to 31.15) GHz (26.5 to 31.15) GHz (31.15 to 50) GHz (31.15 to 50) GHz	0.75 % 2.5 % 0.50 % 4.5 % 1.5 % 6.8 % 1.9 % 2.6 % 6.0 %	E4448A opt 233
Phase Modulation – Mod Rate: (0.2 to 20) kHz 0.3 rad < Dev \leq 0.7 rad Dev > 0.7 rad 0.6 rad < Dev \leq 2.0 rad Dev > 2.0 rad 1.2 rad < Dev \leq 4.0 rad Dev > 4.0 rad 1.3 rad < Dev \leq 4.0 rad Dev > 4.0 rad 2.4 rad < Dev \leq 8.0 rad Dev > 8.0 rad	100 kHz to 6.6 GHz (6.6 to 13.2) GHz (13.2 to 26.5) GHz (26.5 to 31.5) GHz (31.5 to 50) GHz	3.0 % 1.0 % 3.0 % 1.0 % 3.0 % 1.0 % 3.0 % 1.0 % 3.0 % 1.0 %	E4448A opt 233
Transmission (S_{12}/S_{21}) – Measure Linear Phase Linear Mag. Linear Phase Linear Mag.	Type-N Connectors 30 kHz to 2 GHz (2 to 6) GHz	(± 0.76 to ± 39) $^\circ$ (± 0.11 to ± 8.2) dB (± 2.1 to ± 15) $^\circ$ (± 0.25 to ± 1.9) dB	Agilent 8753D VNA w/ 85032B calibration kit

Parameter/Range	Frequency	CMC ^{2,5,6} (\pm)	Comments
Reflection (S_{11}/S_{22}) – Measure Linear Phase Linear Mag. Linear Phase Linear Mag.	Type-N Connectors 30 kHz to 2 GHz (2 to 6) GHz	(± 2.4 to ± 13) $^{\circ}$ (± 0.33 to ± 0.36) dB (± 12 to ± 35) $^{\circ}$ (± 1.1 to ± 1.6) dB	Agilent 8753D VNA w/ 85032B calibration kit
Single Side-Band Phase Noise – Measure Noise Floor: -110 dB -110 dB -130 dB -140 dB -150 dB -155 dB -155 dB -155 dB	Carrier: 50 kHz to 26.5 GHz Offset Frequency: 10 Hz 100 Hz 1 kHz 10 kHz 100 kHz 1 MHz 10 MHz 100 MHz	1.5 dB 1.5 dB 1.5 dB 1.5 dB 1.5 dB 1.5 dB 1.5 dB 1.5 dB	Agilent E4448A
Relative Power – Measure (0 to -10) dB (-10 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (-50 to -60) dB (-60 to -70) dB (-70 to -80) dB (-80 to -90) dB (-90 to -100) dB (-100 to -110) dB (-110 to -120) dB	10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz 10 MHz to 26.5 GHz	0.046 dB 0.053 dB 0.080 dB 0.098 dB 0.11 dB 0.12 dB 0.13 dB 0.17 dB 0.18 dB 0.19 dB 0.19 dB 0.21 dB	Agilent 8902A

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
Absolute Power – Measure			Agilent 437B / E4418B:
(-70 to -30) dBm	10 MHz to 18 GHz	2.7 %	8484A, N-type
(-30 to +10) dBm	100 kHz to 4.2 GHz	1.4 %	8482A, N-type
	(4.2 to 18) GHz	1.9 %	8481A, N-type
	(18 to 26.5) GHz	2.4 %	8485A, 3.5 mm
(+10 to +20) dBm	100 kHz to 4.2 GHz	3.3 %	8482A, N-type
	(4.2 to 18) GHz	3.5 %	8481A, N-type
	(18 to 26.5) GHz	3.8 %	8485A, 3.5 mm

IV. Mechanical

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
Accelerometers ³ – Frequency Response	100 Hz (ref)	1.5 %	Vibration transducer calibration system referenced @ 1 g
	159 Hz (ref)	1.7 %	
	(5 to 100) Hz	2.9 %	
	100 Hz to 1 kHz	2.0 %	
	(1 to 10) kHz	2.5 %	
Scales & Balance ³	Up to 200 g	0.0005 %	Class 1 weights Class 3 weights Class 6, F weights
	Up to 4.5 kg	0.002 %	
	Up to 700 lb	0.01 %	
Force Gages ³	Up to 1000 lbf	0.025 %	Class 6 and F weights
Torque Tools ³	32 in·ozf to 250 ft·lbf	0.65 %	Torque transducers

Parameter/Range	Frequency	CMC ^{2,5,6} (±)	Comments
Pressure Measuring Equipment ³	(0 to 1000) psi	0.28 psi	Additel ADT672
	(≥ 1000 to 10,000) psi	0.28 %	Additel ADT672
	(0 to 1) psid	0.01 % + 0.0015 psid	Process calibrator w/ 700P22
	(-15 to 30) psig	0.01 % + 0.023 psig	700PD5
	(30 to 100) psig	0.01 % + 0.05 psig	700P06
	(100 to 500) psig	0.01 % + 0.25 psig	700P07
	(500 to 1000) psig	0.01 % + 0.50 psig	700P08
(1000 to 3000) psig	0.035 % + 2.5 psig	700P29	
(3000 to 10,000) psig	0.035 % + 8.3 psig	700P31	

V. Thermodynamics

Parameter/Equipment	Range	CMC ^{2,6} (±)	Comments
Relative Humidity ³ – Measure	(10 to 90) % RH	1.1 % RH	Vaisala 141/HMP46
Temperature ³ – Measure	(-20 to 60) °C	0.23 °C	Vaisala 141/HMP46
	(-195 to -38) °C	0.28 °C	Burns 12001 PRT w/ Fluke 2180A
	(-38 to 0) °C	0.13 °C	
	(0 to 200) °C	0.12 °C	
	(200 to 400) °C	0.34 °C	
(400 to 433) °C	0.37 °C		

VI. Time & Frequency

Parameter/Range	Frequency	CMC ^{2,6} (±)	Comments
Frequency – Measuring Equipment	1 mHz to 50 GHz	2.5 parts in 10 ¹² + 0.1 mHz	GPS receiver w/ generator
Frequency – Measure	0.001 Hz to 46 GHz	2.5 parts in 10 ¹² + 0.1 mHz	Counter locked to 10 MHz reference

- ¹ This laboratory offers commercial calibration service and field calibration service.
- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- ⁴ The stated measured values are determined using the indicated instruments (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.
- ⁵ CMC components that can be reasonably attributed to the Unit Under Test have not been utilized in the calculation of the CMC value of this measurement parameter.
- ⁶ In the statement of CMC, percentages are to be read as percent of reading unless indicated otherwise.



Accredited Laboratory

A2LA has accredited

GENERAL CALIBRATION, A TRESICAL COMPANY

Boonton, NJ

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements for ANSI/NCSLI Z540-1-1994 and ANSI/NCSLI Z540.3-2006 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 1st day of March 2016.



A handwritten signature in black ink, appearing to read "J. C. Bunt", written over a horizontal line.

Senior Director Quality & Communication
For the Accreditation Council
Certificate Number 2353.01
Valid to January 31, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.