



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

KING MECHANICAL SPECIALTY  
406 Third Street  
Newburgh, IN 47630  
Daniel Evans Phone: 812 853 8301

CALIBRATION

Valid To: September 30, 2019

Certificate Number: 3125.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
DC Current – Measure	(4 to 20) mA	0.034 mA	Fluke 45 DVM

II. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Pressure – Pneumatic Gauge, Negative Gauge	(0 to 28) inHg	0.0083 inHg	Heise PTE-1 and HQS modules
	(0 to 5) psig	0.0021 psig	
	(0 to 30) psig	0.022 psig	
	(0 to 100) psig	0.039 psig	
	(0 to 300) psig	0.12 psig	
	(0 to 1000) psig	0.4 psig	
(0 to 5000) psig	1.9 psig		

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Torque Wrenches	(20 to 200) in·lbf (5 to 50) in·lbf (10 to 100) ft·lbf (25 to 250) ft·lbf	1.2 % 1.4 % 1.2 % 1.2 %	CDI torque analyzer and transducers

<sup>1</sup> This laboratory offers commercial calibration service

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC, percentages are to be read as percent of reading unless otherwise noted.



## Accredited Laboratory

A2LA has accredited

### KING MECHANICAL SPECIALTY

*Newburgh, IN*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 3<sup>rd</sup> day of October 2017.

A handwritten signature in black ink, written over a horizontal line.

President and CEO  
For the Accreditation Council  
Certificate Number 3125.01  
Valid to September 30, 2019

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*