



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

VIBRATION RESEARCH CORPORATION
1294 Chicago Drive
Jenison, MI 49428
Jade Vande Kamp Phone: 616 669 3028

CALIBRATION

Valid To: October 31, 2019

Certificate Number: 3515.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
DC Voltage – Measure	1V 10V	0.00018 % 0.0043 %	Keysight 34470A

Parameter/Range	Frequency	CMC ^{2,3} (±)	Comments
AC Voltage – Measure 10V 1V	(1 to 5) Hz (5 to 10) Hz 10 Hz to 20 kHz (20 to 50) kHz	0.13 % 0.14 % 0.09 % 0.12 %	Keysight 34470A

II. Time and Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
Frequency – Measure	(3 to 40) Hz (40 to 100) Hz 100 Hz to 1kHz (1 to 50) kHz	810 µHz/Hz 350 µHz/Hz 82 µHz/Hz 81 µHz/Hz	Keysight 34470A

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, the value is defined as the percentage of range.



Accredited Laboratory

A2LA has accredited

VIBRATION RESEARCH CORPORATION

Jenison, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 6th day of October 2017

A handwritten signature in black ink, appearing to read "L. J. ...", positioned above a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 3515.01
Valid to October 31, 2019

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.