



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

STANDRIDGE GRANITE CORP.
9437 Santa Fe Springs Road
Santa Fe Springs, CA 90670
Flor Espinoza-Carrales Phone: 562 946 6334

CALIBRATION


Valid To: March 31, 2020

Certificate Number: 3559.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Granite Surface Plates ³ –			E.G., granite, cast iron, ceramic, steel
Flatness	Up to 480 in	23 µin	Autocollimator
Repeat Reading*	Up to 480 in	14 µin	Repeat-o-meter, (0.000020 indicator) *Only valid in connection with flatness calibration

(A2LA Cert. No. 3559.01) 04/30/2018 

Page 1 of 2

Parameter/Equipment	Range	CMC ^{2, 4, 5} (\pm)	Comments
Accessories –			E.G., granite, cast iron, ceramic, steel
Straightness	Up to 144 in	16 μ in	Autocollimator
Perpendicularity	Up to 144 in	(23 + 0.40L) μ in	Autocollimator with penta prism
Parallelism	Up to 48 in	28 μ in	Electronic amplifier
	Up to 168 in	31 μ in	Autocollimator
Length	Up to 12 in	(38 + 3.0L) μ in	Gage blocks

¹ This laboratory offers commercial and field calibration services and new manufactured products.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately a 95 % level of confidence, usually using a coverage factor of $k=2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the uncertainties achievable on a customer's site can normally be expected to be larger than the CMC that the accredited laboratory has been assigned on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the CMC.

⁴ In the statement of CMC, the numerical value L is for the nominal length of the device measured in inches.

⁵ The CMC values listed do NOT include long- term stability of the UUT, e.g., for Surface Plates the tolerances are not included.



Accredited Laboratory

A2LA has accredited

STANDRIDGE GRANITE CORP.

Santa Fe Springs, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 30th day of April 2018.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 3559.01
Valid to March 31, 2020

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.